CERTIFICAT

UCB_F_12.02 2012-02



CERTIFICATE

No. Z1 13 08 14544 100

Holder of Certificate: Fluke Corporation

> 6920 Seaway Boulevard Everett, WA 98203-9090

USA

Production Facility(ies):

47951

Certification Mark:



Product: Multimeters

(Digital Multimeter)

Model(s): 106, 107

Parameters: 3 VDC Rated Input Voltage:

(2 x 1.5V AAA batteries IEC LR03)

Measuring Ratings: 600V / CAT III **Protection Class:** \parallel

IP Protection: **IP40** 0°C to +40°C Temperature Ambient:

Tested according to:

EN 61010-1:2010 EN 61010-2-030:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

PI1307864-000

Date, 2013-08-15

Page 1 of 1



