



Certificate of Compliance

Certificate: 2501923

Master Contract: 162029

Project: 2510281

Date Issued: March 23, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: Mike Meisner

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, ASCT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Multimeter, Model 28IIEx; battery powered by 3 AAA batteries, NEDA 24A IEC LR3

Notes:

1. The above model is Equipment Class II, Pollution Degree 2, Measurement Category: CAT III 1000V/CAT IV 600V.
2. With Accessory Test Leads Model TL175.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements + CAN/CSA C22.2 No 61010-1-04 Update No. 1:2008

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements